

CERTIFICATE

It is hereby confirmed that the

Fume cupboard controller FC200

of the manufacturer

Hans Schneider Elektronik GmbH

consisting of

Fume cupboard controller FC200

Measuring system with damper and fast actuator VK-250-PPS-K-0-FF

Sash position sensor SPS 100

Airflow sensor AFS100

was tested according to

EN 14175-6:2006

within a duct pressure range of **100 – 600 Pa**.

The details and data of the complete test procedure
are content of the test report

FCS/EN/82/T6/2025

which was given to the manufacturer on


27.03.2025.

Tintschl BioEnergie und Strömungstechnik AG
Goerdelerstr. 21
91058 Erlangen
Germany

Tested and checked by:



Alexander Strobel, M. Sc.



Manzur Ahmed, M. Sc.

Erlangen, 27.03.2025



BioEnergie und Strömungstechnik AG